

Importance of STEC Determination

Shiga toxin-producing *Escherichia coli* strains (non-O157 STEC) have become an increasing public health concern. Some of the non-O157 STEC possess the same range of virulence factors as *E. coli* O157:H7, including the locus of enterocyte effacement and production of Shiga toxin. STEC has been implicated in numerous outbreaks, causing serious illness (hemolytic uremic syndrome), or death.

A study from the CDC showed that from 1982 to 2002 approximately 70% of non-O157 STEC infections in the USA were caused by strains from one of six major serogroups: O104, O45, O103, O111, O121, and O145. Non O-157 STEC has been found in ground beef and in cattle hides, and in feces at levels comparable to those of *E. coli* O157. Bovine feces can be a source of environmental contamination (soil, water) which can lead to secondary contamination of produce growing in fields.

E. coli **O104:H4** is an entero-aggregative strain and is the agent of the 2011 European outbreak that caused 48 deaths and 3,785 cases. The “O” identifies the cell wall lipopolysaccharide antigen, and the “H” identifies the flagella antigen. This strain of *E. coli* is a novel strain that has acquired the Shiga toxin genes presumably by horizontal gene transfer. The strain is characterized by the following genetic markers: ● Shiga toxin stx2 positive ● terE positive (telluride resistance gene cluster) ● eae negative (intimin adherence gene) ● β-lactamases ampC, ampD, ampE, ampG, ampH are present.

It is difficult to distinguish pathogenic non-O157 STEC strains from non-pathogenic *E. coli* strains because the former rarely possess any distinguishing phenotypic or biochemical characteristics from the latter. Therefore, methods such as the latex agglutination test described in this User's Guide have been developed by the USDA-Agricultural Research Service Eastern Regional Research Center (USDA-ARS-ERRC) to help on the identification of these STEC strains. This latex agglutination method is part of the testing protocol utilized and mandated by the FSIS for testing ground beef and beef trim and described in the USDA-FSIS Microbiology Laboratory Guidebook (MLG) Chapter 5B “Detection and Isolation of non-O157 Shiga-Toxin Producing *Escherichia coli* Strains (STEC) from Meat Products and Carcass and Environmental Sponges”.

General Limited Warranty: Abraxis, Inc. warrants the products manufactured by the Company, against defects and workmanship when used in accordance with the applicable instructions for a period not to extend beyond the product's printed expiration date. **Abraxis, Inc. makes no other warranty, expressed or implied. There is no warranty of merchantability or fitness for a particular purpose.**

For ordering or technical assistance contact:

Abraxis, Inc.
124 Railroad Drive
Warminster, PA 18974
Tel.: (215) 357-3911
Fax: (215) 357-5232
Email: info@abraxiskits.com
WEB: www.abraxiskits.com

R103018

E. coli O104:H4 Latex Test Kit

Latex Agglutination Test for the Detection of *E. coli* O104:H4



Product No. 541060 (50 Test)

1. General Description

The Abraxis *E. coli* O104:H4 Test is a rapid latex agglutination test, designed solely for the presumptive identification of *Escherichia coli* serogroup O104 cultured on TSA agar plate. The Abraxis *E. coli* O104:H4 Latex Test Kits should be used as part of the USDA-FSIS test protocol described in the USDA-FSIS Microbiology Laboratory Guidebook (MLG) Chapter 5B “Detection and Isolation of non-O157 Shiga-Toxin Producing *Escherichia coli* Strains (STEC) from Meat Products and Carcass and Environmental Sponges”.

2. Safety Instructions

Biological waste should be decontaminated by autoclaving or by using another effective method. Discard samples according to local, state and federal regulations.

3. Storage and Stability

The *E. coli* O104:H4 Latex Test Kit should be stored between 4–8°C (**do not freeze**) until the expiration date as shown on the label. All reagents and samples to be analyzed should be at room temperature before use.

4. Test Principle

The polystyrene latex particles provided in the kit are coupled to antibodies against *E. coli* serotype O104 (according to Medina et al). When the latex particles are mixed on a test card with fresh colonies of *E. coli* O104, the bacteria will bind to the antibody causing the latex particles to agglutinate (positive reaction). Bacteria that are not *E. coli* O104 will not bind to the antibody and will not agglutinate the latex particles (negative reaction).

5. Limitations of the Test

If a positive result is obtained on an unknown organism, further test such as PCR should be carried out for confirmation. Apply good judgment to any test result, particularly when preliminary positive results are observed.

6. Warning and Precautions

- This product is for *in vitro* diagnostics use only.
- Do not freeze reagents.
- Do not allow reagents to become contaminated by using dirty transfer pipettes.
- Use reasonable judgment when interpreting the test results.
- Prior to use, ensure that the product has not expired by verifying that the date of use is prior to the expiration date on the label.
- Avoid cross-contamination of samples by using a new sample stick for each sample.
- Use only the *E. coli* test reagents from one kit lot** (do not mix with other lots), as they have been adjusted in combination.
- Specimens may contain pathogenic organisms, handle with appropriate precautions.
- Ensure that reagent bottle caps are tight after each use to prevent drying of reagents.
- Reagents contain 0.05-0.1% sodium azide as a preservative. Sodium azide may react with lead or copper plumbing to produce metal azides which might cause explosion. To prevent azide accumulation in plumbing, flush with copious amounts of water immediately after disposal.

7. Sample Collection and Handling

Colonial growth removed from the agar surface of modified Rainbow Agar (mRBA) or tryptic soy agar with 5% sheep blood (SBA) plates is best suited for this testing procedure.

8. Control Procedures

The control reagents provided should be used to check the correct working of the latex reagents each day before routine tests are performed.

The Positive Control suspension must cause visible agglutination with the Antibody-Latex Reagent within one (1) minute.

The Negative Control Suspension should cause no agglutination within one (1) minute.

Do not use the test kit if reactions with the control suspensions are incorrect.

A. Materials Provided

1. Negative Control. Suspension of inactivated *E. coli* cells (K12) in buffer, 1 vial, 0.5 mL.
2. Positive Control. Suspension of inactivated *E. coli* O104:H4 cells in buffer, 1 vial, 0.5 mL.
3. PBS (1X), 1 vial, 1.0 mL.
4. Antibody-Latex Beads Reagent. A suspension of red or blue latex particles coupled to specific rabbit IgG to *E. coli* O104:H4 serotype, 1.5 mL vial with pink cap (enough for 50 tests).
5. Control Latex Reagent. A suspension of red or blue latex particles sensitized with pre-immune rabbit globulin, 1 vial, 1.0 mL vial with white cap.
6. Test Cards. Disposable reaction cards, 5 cards (10 reactions each).
7. Sample Mixing Sticks (100).
8. Transfer Pipettes (20). Color coded (10 green, 5 blue, 5 red)
9. User's guide.

B. Additional Materials (not provided with the test)

1. Agar Plates.
2. Disinfectant Solution e.g. Sodium hypochloride solution >1.3% w/w.

C. Test Preparation

1. All reagents and samples to be analyzed should be at room temperature before use.
2. Thoroughly suspend the latex reagent and controls by agitation.

D. Assay Procedure

1. Bring all reagents and samples to room temperature. Make sure the latex suspensions and control are well mixed by vigorous shaking.
2. Using one of the provided transfer pipettes (red), place one drop of PBS onto one (1) circle on the test card. If more than one sample is being tested use additional circles on the test card.
3. Using one of the sample mixing sticks (or an inoculating loop), pick a portion of a suspect colony from the agar plate and thoroughly emulsify in the drop of PBS of one of the circles.
4. Using one of the transfer pipettes (red) add one (1) drop of the Positive Control to a second circle
5. Using another transfer pipette (blue), add one (1) drop of the Negative Control to a third circle.
6. Dispense one (1) free falling drop (with vial held vertically) of the *E. coli* O104 Latex-Antibody bead reagent onto each circle (Positive, Negative, and sample(s))
7. Rotate the test card using a complete circular motion (through 3 planes) for up to one (1) minute or until agglutination is evident, whichever occurs first. Record the results.
8. If agglutination with the test reagent does occur, it is necessary to test a further portion of the colony with the Control Latex Reagent to ensure that the isolate is not an auto-agglutinating strain.

NOTE: All mixing sticks, cards, etc. should be disposed in disinfectant or autoclave waste containers.

E. Interpretation of Results

Agglutination of the test latex within one (1) minute is a positive result. This indicates the presence of *E. coli* serogroup O104:H4.

No agglutination occurring within one (1) minute is a negative result. This indicates the absence of *E. coli* serogroup O104:H4.

NOTE: Some strains of *E. coli* are difficult to emulsify in saline and may give a stringy type reaction with the test reagents. This does not look like true agglutination and should be ignored. If this stringiness is found to be too severe for a correct judgment to be made then the colony should be suspended in 1-2 drops of PBS. Allowing the lumps to settle and re-test.

If a positive result is obtained on an unknown organism, further test such as PCR should be carried out for confirmation. Apply good judgment to any test result, particularly when preliminary positive results are observed.

F. Cross-reactivity Profile

<i>E. Coli</i>	
Serotype	Agglutination
O104:H4	“+++”
O26	“-“
O45	“-“
O103	“-“
O111	“-“
O145	“-“
O157	“-“
K12	“-“

“+ + +” = Strong positive agglutination

“- “ = Negative agglutination

G. Additional Analysis

Positive samples must be confirmed as described in the USDA-FSIS test protocol described in the USDA-FSIS Microbiology Laboratory Guidebook (MLG) Chapter 5B “Detection and Isolation of non-O157 Shiga-Toxin Producing *Escherichia coli* Strains (STEC) from Meat Products and Carcass and Environmental Sponges”.

H. References

(1) Medina, M., Shelver, W., Fratamico, P., Fortis, L., Narang, N., Cray, W. Jr., Esteban, E., Tillman, G., and Debroy, C. Latex agglutination assays for detection of Non-O157 Shiga Toxin-Producing *Escherichia coli* Serogroups O26, O45, O103, O111, O121 and O145. Journal of Food Protection 75(5):819-826.

(2) USDA-FSIS Microbiology Laboratory Guidebook (MLG) Chapter 5B “Detection and Isolation of non-O157 Shiga-Toxin Producing *Escherichia coli* Strains (STEC) from Meat Products and Carcass and Environmental Sponges”.